

# **BIOCOMPATIBILITY CERTIFICATE**

**Testmaterial:** PA 2200

**Supplier:** EOS GmbH  
Pasinger Strasse 2, D-82152 Planegg

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the product PA 2200 according to ISO 10993-1:

**CYTOTOXICITY**

**SENSITISATION, polar extract**

**SENSITISATION, non-polar extract**

**INTRACUTANEOUS REACTIVITY**

**Results:** The product did not show any adverse effects in the studies performed. Therefore, the biocompatibility of the test material was proved.

**BSL BIOSERVICE Scientific Laboratories GmbH Munich**

Behringstraße 6

D-82152 Planegg



Dr. Achim Albrecht

Biological Safety Testing

Date: April 10, 2001

